

केन्द्रीय मुहर विभाग -2

संदर्भ - : के मू वी-2/16:17423

20 09 2022

विषय: IS 17423:2021 के amendment no. 1 के अनुपालन के दिशा निर्देश

यह उपरोक्त विषय के संदर्भ में है।

सक्षम अधिकारी द्वारा अनुमोदित दिशा निर्देश अनुपालन हेतु संलग्न है।

सभी क्षेत्रीय/शाखा कार्यालयों से अनुरोध है की दिशा निर्देशों का तत्काल प्रभाव से अनुपालन सुनिश्चित करें।

आदित्य दास
वैज्ञानिक D

प्रमुख (के मू वी 2)

सभी क्षेत्रीय/शाखा कार्यालय/प्रयोगशालाएँ/TXD/LRMD

CENTRAL MARKS DEPARTMENT-2

OurRef:CMD-2/16:17423

20 09 2022

Subject: Guidelines for implementation of amendment no. 1 to IS 17423:2021

This has reference to the subject mentioned above.

The Competent Authority has approved the enclosed Guidelines for implementation.

All ROs/BOs are requested to ensure the implementation of the above Guidelines with immediate effect.

(Aditya Das)
Scientist D

Head (CMD-2)

All ROs/BOs/Labs/TXD/LRMD

CENTRAL MARKS DEPARTMENT-2

OurRef:CMD-2/16:17423

Date: 20 09 2022

Subject: Guidelines for implementation of Amendment no. 1 to IS 17423:2021 – (Medical Textiles – Bio-Protective Coveralls – Specification)

1. Amendment no. 1 to IS 17423:2021 has been published.
2. This amendment has been issued to correct the limit specified for the requirement of Breathability (water vapour transmission rate) specified under Table 1 Requirements for Bio-protective Coveralls from “Max” to “Min”.
3. IS 17423:2020 is currently under revision to IS 17423:2021. The (extended) last date of implementation of the revised standard is 12 Jan 2023 as already informed vide CMD-2’s circular dated 20 Jul 2022. CMD-2 has also circulated the guidelines for implementation of the revised standard on 17 Nov 2021.
4. Since this amendment is a correction to the standard, BOs shall inform the Applicants and Licensees under their jurisdiction about implementation of the above amendment within a week of issuance of these guidelines.
5. BOs shall also advise them to ensure that the evidence of conformity (test reports) to the revised standard to be submitted by them by 12 Jan 2023, as per above circulars dated 17 Nov 2021 and 20 Jul 2022, should indicate conformity of the product to the standard including amendment no. 1.

The evidences of conformity (test reports) if any already submitted by manufacturers for implementation of IS 17423:2021 shall also be reviewed to ensure that they meet the standard considering amendment no. 1.

6. The above guidelines come into force with immediate effect.

AdityaDas
Sc.D

Head (CMD-2)
DDG (Certification)

AMENDMENT NO. 1 AUGUST 2022
TO
IS 17423 : 2021 MEDICAL TEXTILES — BIO-PROTECTIVE COVERALLS —
SPECIFICATION
(First Revision)

[Page 3, Table 1, clause 5.2, Sl No. (iii)] — Substitute 'Min' for 'Max'

(TXD 36)

Publication Unit, BIS, New Delhi, India