

BUREAU OF INDIAN STANDARDS

DRAFT FOR COMMENTS ONLY

(Not to be reproduced without permission of BIS or used as an Indian Standard)

भारतीय मानक मसौदा
दंत चिकित्सा - एंडोडॉन्टिक उपकरण
भाग 6 संख्यात्मक कोडिंग प्रणाली

Draft Indian Standard
Dentistry — Endodontic Instruments
Part 6 Numeric Coding System
ICS 11.060.20

Dentistry Sectional Committee, MHD 08

Last date for comments: 13 April 2024

NATIONAL FOREWORD

(Adoption clause will be added later)

The text of ISO Standard has been approved as suitable for publication as an Indian Standard without deviations. Certain conventions are however not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- Wherever the words 'International Standard' appear referring to this standard, they should be read as 'Indian Standard'
- Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

In this adopted standard, reference appears to certain International Standards for which Indian Standards also exist. The corresponding Indian Standards which are to be substituted in their respective places are listed below along with their degree of equivalence for the editions indicated:

<i>International Standard</i>	<i>Corresponding Indian Standard</i>	<i>Degree of Equivalence</i>
ISO 1942, Dentistry — Vocabulary	IS 17895 : 2023/ISO 1942 : 2020, Dentistry Vocabulary	Identical
ISO 3630-1, Dentistry — Endodontic instruments — Part 1: General requirements	IS 17859 (Part 1) : 2022/ISO 3630-1:2019, Dentistry Endodontic instruments Part 1: General requirements	Identical

Scope

This document specifies a coding system for specific characteristics of endodontic instruments, with a 12-digit code identifying general and specific characteristics of instruments or groups of instruments. The numerals identify the type of instrument, nominal size, taper, length, working part material, type and material of handle or shank.

The technical content of the document has not been enclosed as it is identical with the corresponding ISO standard. For details, please refer to ISO/TR 3630-6:2023 or kindly contact:

Head
Medical Equipment and Hospital Planning Department
Bureau of Indian Standards
9 Bahadur Shah Zafar Marg
New Delhi-110002
Email: mhd@bis.gov.in
hmhd@bis.gov.in