

(PREVIEW)

Indian Standard

MEDICAL ELECTRICAL EQUIPMENT

PART 2 PARTICULAR REQUIREMENTS FOR THE SAFETY

Section 5 Ultrasonic Physiotherapy Equipment

SECTION ONE – GENERAL

1.1 Scope

Addition

This Particular Standard specifies the requirements for safety of ULTRASONIC PHYSIOTHERAPY EQUIPMENT used in medical practice, as defined in 2.1.101.

This Particular Standard does not apply to:

- EQUIPMENT in which a tool is driven by ULTRASOUND (for example EQUIPMENT used in surgery or dentistry);
- EQUIPMENT in which focused ULTRASOUND pulse waves are used to destroy conglomerates such as stones in the kidneys or the bladder (lithotrites) (for information refer to IEC 60601-2-36);
- ULTRASONIC PHYSIOTHERAPY EQUIPMENT in which focused ULTRASOUND pulse waves are used.