

**IS : 5185 -1969**

# **Indian Standard**

## **SPECIFICATION FOR ARTICULATORS, DENTAL**

### **0. FOREWORD**

**0.1** This Indian Standard was adopted by the Indian Standards Institution on 16 July 1969, after the draft finalized by the Dental Instruments Sectional Committee had been approved by the Consumer Products Division Council.

**0.2** This standard is one of a series of Indian Standards on dental instruments and has been formulated at the instance of the Advisory Committee for Development of Surgical Instruments, Equipment and Appliances, Government of India. Other specifications published so far in the series are given on p 10.

**0.3** For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS : 2-1960\*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

### **1. SCOPE**

**1.1** This standard covers the requirements of plane-line articulators, Types 1 and 2 and free-plane articulator used for setting the artificial teeth in proper relation to each other in denture conduction and for other purposes, where upper and lower casts have to be articulated.