

60

SYNOPSIS

Doc: MHD 06 (14327) IS/ISO 8637-2:2018, Extracorporeal systems for blood purification - Part 2: Extracorporeal blood circuit for haemodialysers, haemodiafilters and haemofilters

Scope:

This standard specifies requirements for the blood circuit for devices used in extracorporeal blood filtration therapies such as, but not limited to, haemodialysis, haemodiafiltration, haemofiltration and transducer protectors (integral and non-integral) intended for use in such circuits.

Salient features of content

This standard does not apply to:

- haemodialysers, haemodiafilters or haemofilters;
- plasmafilters;
- haemoperfusion devices;
- vascular access devices;
- blood pumps;
- pressure monitors for the extracorporeal blood circuit;
- air detection devices;
- systems to prepare, maintain or monitor dialysis fluid;
- systems or equipment intended to perform haemodialysis, haemodiafiltration, haemofiltration or haemoconcentration.

c) Type/grades/classes, if any covered in the standard: Nil.
